

## Certificate of Analysis

Publication No: 025CA070V0110

### **Spike-In RNA Variant Mixes**

Cat. No: 025.03

Kit Lot 216652830

Size: 3 x 4 µl

Concentration: 25 ng/µl

Expiration date: 30 Sept 2016

### **Kit components tested:**

1. SIRV Mix E0: Lot 216051530
2. SIRV Mix E1: Lot 216151530
3. SIRV Mix E2: Lot 216251530

### **Functional Testing:**

- Lot confirms to properties described in User Guide 025UG063V0110, with one exemption:
- SIRV502 is fragmented.
- Lot passes testing for RNase Contamination:
  - No change in microfluidics trace characteristics was observed when the mixes were incubated at 37°C for 14 hours.
  - Testing with RNaseAlert (Integrated DNA Technologies, Inc.) was negative for RNase presence in mix solutions.

**Result: Pass**



*Quality Control*

---

FOR RESEARCH USE ONLY. NOT INTENDED FOR DIAGNOSTIC OR THERAPEUTIC USE.

---